



# CALIBRATION REPORT

ORDER No. | }

MAY 31, 2016

PAGE 1 OF 1

MANUFACTURER: OHM-LABS  
 DESCRIPTION: CURRENT SHUNT  
 MODEL: CSW-100  
 SERIAL:

PROCEDURE: CS CAL  
 LAB ENVIRONMENT: 23.7 °C / 39 %RH  
 CALIBRATION DATE: 31/MAY/2016  
 CALIBRATION DUE

MEASUREMENT DATA – AS FOUND / AS LEFT		
APPLIED CURRENT	MEASURED VALUE	UNCERTAINTY
20 A	10.000 34 mΩ	9 μΩ/Ω
40	10.000 44	9
60	10.000 48	8
80	9.999 42	11
100	9.999 87	7

NOTES:  
 SHUNT WAS ALLOWED TO FULLY STABILIZE AT EACH APPLIED CURRENT.

<u>STANDARDS USED</u>			
ID	DESCRIPTION	MAKE & MODEL	CAL DUE
AS3012	RESISTANCE STANDARD	OHM-LABS 201	30/SEP/2016
AS3403	RESISTANCE BRIDGE	GUILDLINE 9975	28/FEB/2017
AS3407	RANGE EXTENDER	GUILDLINE 9923	28/FEB/2017

COMMENTS:

OHM-LABS, INC. CERTIFIES THAT THIS CALIBRATION IS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST), OR ANOTHER RECOGNIZED NATIONAL MEASUREMENT INSTITUTE, OR DERIVED BY A RATIO TYPE SELF-CALIBRATION TECHNIQUE, AND IS ACCREDITED TO ISO/IEC 17025. OHM-LABS' QUALITY CONTROL SYSTEM MEETS THE REQUIREMENTS OF ANSI/NC SL Z540-1-1994. THE REPORTED UNCERTAINTIES REPRESENT EXPANDED UNCERTAINTIES EXPRESSED AT A CONFIDENCE LEVEL OF APPROXIMATELY 95 %, USING A COVERAGE FACTOR OF K=2. THIS UNCERTAINTY IS AT THE TIME OF TEST ONLY AND DOES NOT TAKE INTO ACCOUNT TRANSIT, USAGE, DRIFT OVER TIME, OR OTHER FACTORS AFFECTING STABILITY. THIS DOCUMENT CERTIFIES THAT THE ITEMS IDENTIFIED HEREIN COMPLY WITH ALL REQUIREMENTS OF THE ABOVE PURCHASE ORDER, AND THAT THE CALIBRATION PERFORMED WAS IN ACCORDANCE WITH THE CURRENT REVISION LEVEL OF OHM-LABS' QUALITY CONTROL SYSTEM. TRAINED AND QUALIFIED PERSONNEL PERFORMED THE CALIBRATIONS IN ACCORDANCE WITH THE REQUIREMENTS OF ISO/IEC 17025. THIS CERTIFICATE SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT WRITTEN PERMISSION BY OHM-LABS, INC.

PERFORMED BY:

REVIEWED BY:

